



ISTITUTO DI RICERCHE E COLLAUDI

M. MASINI S.r.l.

Sede amministrativa e laboratori: Via Moscova, 11 - 20017 RHO (MI) - ITALIA - Sede Legale: Via S. Eufemia, 2 - 20122 MILANO - CF e P.IVA 00862210150
Tel. +39.02.93.01.517 r.a. - Fax +39.02.93.08.176 - www.istitutomasini.it - istitutomasini@istitutomasini.it

RAPPORTO DI PROVA N. 1576/2-2013

TEST REPORT No. 1576/2-2013

rif. alla norma UNI EN 1865-1:2010 e UNI EN 1789:2010 (pt. 4.5.9 e 5.3)
rif. to standards UNI EN 1865-1:2010 and UNI EN 1789:2010 (pt. 4.5.9 and 5.3)

L'Istituto di Ricerche e Collaudi M. Masini dichiara di avere effettuato le prove del/i prodotto/i sottoindicato/i, seguendo i requisiti indicati nei punti 4.2 (ad esclusione del pt. 4.2.8) e 5.1 della norma UNI EN 1865-1:2010.
The Istituto di Ricerche e Collaudi M. Masini hereby declares that tests have been carried out on the product/s listed hereafter, following the requirements of pt. 4.2 (with the exclusion of pt. 4.2.8) and 5.1 of the standards UNI EN 1865-1:2010.

FABBRICANTE:
Manufacturer

ME.BER. S.r.l. Unipersonale
Strada Langhirano, 270
43124 Fontanini (PR) - ITALY

PRODOTTO "Tipo, identificazione, utilizzo":
Product "Type, designation, use"

**Barella fissa in alluminio montata su carrello
autocaricante mod. "Master" e relativo sistema
di fissaggio all'ambulanza art. 955 PROOF**
*Fixed aluminium stretcher used on an automatic
loading trolley mod. "Master" and support for the
fixing inside the ambulance art. 955 PROOF*

modello model "1515 PROOF"

LUOGO DI PRODUZIONE:
Production site

Strada Langhirano, 270
43124 Fontanini (PR) - ITALY

NORME DI RIFERIMENTO:
Reference standards

UNI EN 1865-1:2010
UNI EN 1789:2010 - pt. 4.5.9 e/and 5.3

ESITO:
Result

POSITIVO - POSITIVE
ad esclusione del pt. 4.2.3 della norma UNI EN 1865-1:2010
except pt. 4.2.3 of the standard UNI EN 1865-1:2010
(dettagli in allegato 1 - *see details in annex 1*)

Data/Date: 20/12/2013

Il Direttore Tecnico - *Technical Manager*
Dr. Ing. Sergio Tosi